## **ARCS Instrument Replacement Form**

Fill out the follo	wing form when	replacing inst	truments at an AR	CS site:				
System: Skyra	d NIP		<u> </u>					
Site: Nauru			_					
OLD Configu	<u>ration</u>							
Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45o)X10-6	Offset	Date of change UTC	Time Start UTC	Time End UTC
NIP	31350E6	Skyrad	26 Sept 2004	8.3338		15 Aug 2005	00:30	01:35
			NIP S/N 31346	L				
NEW Configu	<u>ration</u>		2					
Part Number	Serial Number	Location	Configuration Date	Cal Gain Coefficient (Responsivity at 45o)X10-6	Offset	Copy Config. Logger to PC		
NIP	31346E6	Skyrad	26 Sept 2004	8.1578		yes		
Replacement co	ompleted by: Ob	servers					•	
NOTES:								
Old Part remair	ns on site as spa	re:						
Old part packed	d for shipment to	:						
Other old part o	lisposition (expla	ain): <u>Ren</u>	nains at Nauru for	testing during SVC	0509Ncal			

DW(@PS)000etermined during SV0509Ncal that NIP was filled with condensation.